

Analytical Test Report

CBD Bath Bomb

C&R Oregon

030-1001972282D

Laboratory ID: 1802036-01

Batch ID:

Date Sampled: 02/09/18

Date Printed: 02/15/18

Report cannot be used for OLCC/OHA compliance.

Potency Analysis

Analytical Method: De Backer, Journal of Chromatography b.2009. 11.004 -
SOP 19 & 20

Cannabinoids	mg/g	Notes	
THCA	< LOQ		Total THC < LOQ mg/g
delta 9-THC	< LOQ		
CBGA	< LOQ		
CBDA	< LOQ		Total CBD 0.325 mg/g
CBD	0.325		
CBN	< LOQ		
CBG	< LOQ		
CBC	< LOQ		



Chris Nielsen, Lab Director 2/15/2018

Analytical Test Report

CBD Bath Bomb

C&R Oregon

030-1001972282D

Laboratory ID: 1802036-01

Quality Control Potency

Batch: B18B081 - Potency
Blank(B18B081-BLK1)

Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
delta 9-THC	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
CBGA	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
CBDA	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
CBD	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
CBN	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
CBG	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	
CBC	< LOQ	0.00994	mg/g		02/14/18 18:53	02/14/18 21:29	

LCS(B18B081-BS1)

Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	114	0.00996	mg/g	80-120	02/14/18 18:53	02/14/18 21:14	
delta 9-THC	132	0.00996	mg/g	80-120	02/14/18 18:53	02/14/18 21:14	M, Q2
CBDA	110	0.00996	mg/g	80-120	02/14/18 18:53	02/14/18 21:14	
CBD	97.6	0.00996	mg/g	80-120	02/14/18 18:53	02/14/18 21:14	

Notes and Definitions

- B Analyte detected in method blank, but not associated samples.
 - B2 Analyte detected in sample and associate method blank.
 - C Interference due to co-elution.
 - D Initial result exceeded calibration range, reported data are based on analysis of a dilution.
 - H Non-homogenous sample matrix affecting RPD and/or QC.
 - I Manual Integration was performed.
 - L Duplicate sample relative percent difference (RPD) exceeds QC limits.
 - M Anomalous results due to matrix interference
 - P Peaks manually split.
 - Q1 QC out of limits but still oK
 - Q2 Quality Control outside QC limits. Data considered estimate.
 - Q3 CCV was above the acceptance criteria. Non-detect samples are considered acceptable.
 - Q4 CCV was below the acceptance criteria, however the sample still exceeds the regulatory limit.
 - R Marginal Exceedence.
 - U Reported result is an estimate. The analyte was detected above the calibration range.
 - X Problems with initial analysis, reported data are from reinjection of prepared sample.
- <LOQ - Results below the Limit of Quantitation - Compound not detected



Chris Nielsen, Lab Director 2/15/2018

Page 2 of 2